

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/GJ/26/02)

Certificate of Verification**AXPERT ENTERPRISE, PLOT NO. E 149 + 150 ,GIDC ELECTRONICS ESTATE, SECTOR - 26, GANDHINAGAR, GUJARAT, 382028.-382028, Ph: 9374991015****Certificate No:** IND/GATC/GJ/26/02/26/1**Date of Verification:** 24-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**Axpert Enterprise**,Address-**E 149-150, Electronic Estate, GIDC, Sector - 26, Gandhinagar, Gujarat - 382028, Gandhinagar,Gujarat,382028, Ph:- 9099693822**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ABS - 30	26030001	2026	Class III	30	100 g	5 g	kg	5	6000	7.5

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	25	50 %	230 - 50 Hz	No		

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	No	

Model Approval No(s) - IND/09/14/313**Verification Fee (Rs. + GST) - 2000; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **00**, Dated **24-03-2026**Repaired by / Used by: **AXPERT ENTERPRISE**Next verification falls due on or before: **2027-03-23****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.