

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/RJ/25/03)

Certificate of Verification**Yadav Measurements Private Limited, 373-375, RIICO BHAMA SHAH INDUSTRIAL AREA, KALADWAS, UDAIPUR, Udaipur, Rajasthan, 313001-313003, Ph: 8690222058****Certificate No:** IND/GATC/RJ/25/03/26/178**Date of Verification:** 13-04-2026**I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-HCG DISTRIBUTION PRIVATE LIMITED, Address-HCG DISTRIBUTION PRIVATE LIMITED KHASRA NO.-55/19/1, VILLAGE - DHANKOT, (NEAR SHYAM CHOWK DHANKOT), GURGAON, HARYANA, 122505, IN, Gurugram, Haryana, 122505, Ph:- 9990 595 595**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Other	Secure Meters Limited / Liberty Gas 120	Q1964746	2026	1.5	Natural Gas	2.5 m ³ /h	0.016 m ³ /h	0.25 m ³ /h	20	G-1.6

Unit of Measurement (m ³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m ³	0.001	± 2.0% at Qmax and Qt, ± 3.5% at Qmin	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
23°C to 25°C	23°C to 25°C	935-945 mbar (abs)	< 55%	Yes	702551	Metering FW 23420101 Metering FW CRC 590C2C4A	Yes	Yes	

Model Approval No(s) - IND/09/24/039**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 590 deposited vide Money Receipt No. Wo No. 213, Dated 11-04-2026****Repaired by / Used by: Yadav Measurements Private Limited****Next verification falls due on or before: 2032-04-12****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.