

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/UP/25/11)

Certificate of Verification**CIVA Industrial Solutions, D-177 First Floor, Sector -10, Noida-201301, Ph: 9958050808****Certificate No:** IND/GATC/UP/25/11/26/01**Date of Verification:** 24-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**Lohum Materials Pvt Ltd,Address-B-35, Ecotech 1 Extension, Greater Noida,Gautam Buddha Nagar,Uttar Pradesh,201310, Ph:- 7983578029**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	VELAA / VTT	24021527	2023	III	10KG	20G	1G	GRAM	1G	20g, 50g, 100g, 200g, 500g, 1kg, 2kg, 5kg, 10kg	1G

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	22	60	220	Yes	001	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	0

Model Approval No(s) - IND/09/20/636**Verification Fee (Rs. + GST) - 2360; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 2360** deposited vide Money Receipt No. **1005**, Dated **18-02-2026**Repaired by / Used by: **CIVA Industrial Solutions**Next verification falls due on or before: **2027-02-23****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.