

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/UP/25/11)

**Certificate of Verification****CIVA Industrial Solutions, D-177 First Floor, Sector -10, Noida-201301, Ph: 9958050808****Certificate No:** IND/GATC/UP/25/11/26/10**Date of Verification:** 21-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**SIRIUS FOODS PVT. LTD.,Address-C-17, PHASE-2, INDUSTRIAL AREA, NOIDA,Gautam Buddha Nagar,Uttar Pradesh,201305, Ph:- 9211243003**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	EXCELL/AH-WPS	2502192	2026	III	5kg	10g	0.5g	kg	0.5g	10000	0.75g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	28.1	51	230V-50HZ	Yes	IND/GATC/UP/25/11	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

**Model Approval No(s) - IND/09/2003/143****Verification Fee (Rs. + GST) - 9440; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 9440** deposited vide Money Receipt No. **CIS/26-27/0029**, Dated **21-04-2026**Repaired by / Used by: **CIVA Industrial Solutions**Next verification falls due on or before: **2027-04-20****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**