



# GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/05)

## Certificate of Verification

Government Approved Test Centre

**Green Globe Fuel Solutions Pvt. Ltd., Green Globe Fuel Solutions Pvt. Ltd. Greenhouse I, II & III, Building No. 5 & 6, Radhe Krishna Industrial Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane – 421302, Maharashtra- India-421302, Ph: 9820056837**

**Certificate No:** IND/GATC/MH/25/05/26/01

**Date of Verification:** 07-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**GREENGLOBE FUEL SOLUTIONS PVT LTD, Address-Greenhouse I & II, Building.No.5 & 6, Radhe Krishna Industrial Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane – 421302, Maharashtra, India., Thane, Maharashtra, 421302, Ph:- 7718801149**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	SI (D)
Diaphragm	GRECO	24260105001	2026	1.5	Natural Gas	2.5 scmh	0.016 scmh	0.25 scmh	2

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m3/hr	0.01	1.5%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature 20±5(°C)	Gas Temperature 40±5(°C)	Operating Pressure	Relative Humidity 40±5(%)	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011
25	25	0.5 bar	55	Yes	G10369	4.0.1	Yes

**Model Approval No(s) - TCM/143/23-5931**

**Verification Fee (Rs. + GST) - 590. Carriage / Conveyance, adjusting Charges, etc. (+ GST)- 0**

**Total Rs.590 deposited vide money Receipt. No.123 Dated 07-02-2026 Repaired by/ Used by : Green Globe Fuel Solutions Pvt. Ltd.**

**Next verification falls due on or before: 2031-07-02**

**Signature of Principal Officer**

**Note:**

**1.The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**

