

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/KL/26/04)

Certificate of Verification**Interweighing Pvt. Ltd, 49/470 D1,3rd floor, Asian Tower,Vyttila ,Cochin, Kerala -682019-682019, Ph: 8590601636****Certificate No:** IND/GATC/KL/26/04/26/458**Date of Verification:** 30-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**Interweighing Pvt. Ltd,,Address-49/470 D1,3rd floor, Asian Tower,Vyttila,Ernakulam,Kerala,682019, Ph:-8803333444**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	YESWEIGH	Y09489	2026	III	30kg	100g	5g	kg	5g	6000	7.5g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification on Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	33	75	230V AC	Yes	26/04	NIL

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/20/15**Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. N/A** deposited vide Money Receipt No. **N/A**, Dated **N/A**Repaired by / Used by: **Interweighing Pvt. Ltd**Next verification falls due on or before: **2027-04-29**

Signature of Principal Officer**Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.