

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/RJ/25/06)

Certificate of Verification**Capital Innotech Pvt., E-100, Mewar Industrial Area, Madri-313003, Ph: 9891079083****Certificate No:** IND/GATC/RJ/25/06/26/12938**Date of Verification:** 16-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**Central UP Gas Limited**, Address-**7th Floor UPSIDC Complex, A-1/4, Lakhanpur, Kanpur, Kanpur Nagar, Uttar Pradesh, 208024, Ph:- 5122585001**

| Type of Gas Meter | Manufacturer / Model / Brand / Series | Serial Number | Year of Manufacture | Accuracy Class | Gas Type | Maximum Flow Rate (Qmax) | Minimum Flow Rate (Qmin) | Transitional Flow Rate (Qt) | Size (DN) | G-Value |
|-------------------|---------------------------------------|----------------|---------------------|----------------|-------------|--------------------------|--------------------------|-----------------------------|-----------|---------|
| Diaphragm | CAPITAL G1.6MS | CI-26-02-13256 | 2026 | 1.5 | Natural Gas | 2.5 | 0.016 | 0.25 | 20 | 1.6 |

| Unit of Measurement (m ³ / kg) | Indicating Device Resolution | Maximum Permissible Error (MPE) | Visual Examination | Accuracy Test at Qmin | Accuracy Test at Qt | Accuracy Test at Qmax | Overall Performance Result |
|---|------------------------------|---|--------------------|-----------------------|---------------------|-----------------------|----------------------------|
| m ³ | 0.0001 | Qmax. to Qt ±1.5%, Qt to Qmin. ± 3% | Pass | Pass | Pass | Pass | Pass |

| Ambient Temperature | Gas Temperature | Operating Pressure | Relative Humidity | Verification Seal Affixed | Seal Identification No(s) | Software / Firmware Version & Checksum (if applicable) | Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial / billing transactions | Remarks (if any) |
|---------------------|-----------------|--------------------|-------------------|---------------------------|---------------------------|--|--|---|------------------|
| 24±2 | 24±2 | 94±3kPa (Pabs) | 55±10 | Yes | 602132566 | - | Yes | Yes | - |

Model Approval No(s) - IND/09/22/924**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **Not Applicable**, Dated **16-04-2026**Repaired by / Used by: **Capital Innotech Pvt.**Next verification falls due on or before: **2033-04-15****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.