

**GOVERNMENT APPROVED TEST CENTRE**  
(IND/GATC/TN/25/02)

**Certificate of Verification**

**Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011232**

Certificate No: IND/GATC/TN/25/02/26/494

Date of Verification: 19-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-ISHTAA SCALES INC,Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400

| Type of Instrument | Manufacturer / Model / Brand / Series Designation | Serial Number | Year of Manufacture | Accuracy Class (III) | Maximum Capacity (Max upto 150 kg) | Minimum Capacity (Min) | Verification Scale Interval (e) | Unit of Measurement:kg | Actual Scale Interval (d) | No. of Verification Intervals (n = Max / e) | Maximum Permissible Error (MPE) |
|--------------------|---|---------------|---------------------|----------------------|------------------------------------|------------------------|---------------------------------|------------------------|---------------------------|---|---------------------------------|
| Electronic         | ISHTAA-ITB  | VGCS490       | 2026                | III                  | 15/30kg                            | 40g                    | 2/5g                            | kg                     | 2/5g                      | 6000  | 7.5g                            |

| Visual Examination | Zero Setting / Zero Tracking Test | Eccentricity Test | Repeatability Test | Accuracy / Weighing Performance Test | Tare Device Test | Overall Verification Result | Ambient Temperature | Relative Humidity | Supply Voltage (if electronic) | Verification on Seal Affixed | Seal Identification No(s) | Software Identification / Version /Checksum (if applicable) |
|--------------------|-----------------------------------|-------------------|--------------------|--------------------------------------|------------------|-----------------------------|---------------------|-------------------|--------------------------------|------------------------------|---------------------------|---|
| Pass               | Pass                              | Pass              | Pass               | Pass                                 | Pass             | Pass                        | 31.7                | 71                | 230V-50HZ                      | Yes                          | IND/GATC/TN/25/02         |   |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial transactions | Remarks (if any): |
|--|---|-------------------|
| Yes  | Yes   |                   |

Model Approval No(s) - IND/09/09/84

Verification Fee (Rs. + GST) - 7080; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. 164, Dated 19-03-2026

Repaired by / Used by: Ishtaa Scales Inc.

Next verification falls due on or before: 2027-03-18

*V. B. B.*

Signature of Principal Officer

**Note:**

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC  
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

