

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/25/05)

**Certificate of Verification**

Government Approved Test Centre

**Green Globe Fuel Solutions Pvt. Ltd., Green Globe Fuel Solutions Pvt. Ltd. Greenhouse I, II & III, Building No. 5 & 6, Radhe krishna Industrial Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane – 421302, Maharashtra- India-421302, Ph: 9820056837**

**Certificate No:** IND/GATC/MH/25/05/26/03**Date of Verification:** 11-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-**SABARMATI GAS**,Address-**Plot No. 907, Sector - 21,,Gandhinagar,Gujarat,382021, Ph:- 7966737600**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	GRECO	24260108000	2026	1.5	Natural Gas	2.5 m3/hr	0.016 m3/hr	0.25 m3/hr	20	G1.6

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m3/hr	0.01	1.5%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
20	41	0.5 bar	55	Yes	123456	4.0.1	Yes	Yes	

**Model Approval No(s) - TCM 143/23-5931****Verification Fee (Rs. + GST) - 590. Carriage / Conveyance,adjusting Charges,etc. (+ GST)- 590**

**Total Rs.590 deposited vide money Receipt. No.123456789 Dated 11-02-2026 Repaired by/ Used by : Green Globe Fuel Solutions Pvt. Ltd. Next verification falls due on or before: 2033-02-11**

**Signature of Principal Officer****Note:****1.The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**

