

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/05)

Certificate of Verification**Green Globe Fuel Solutions Pvt. Ltd., Green Globe Fuel Solutions Pvt. Ltd. Greenhouse I, II & III, Building No. 5 & 6, Radhe Krishna Industrial Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane - 421302, Maharashtra- India-421302, Ph: 9820056837****Certificate No:** IND/GATC/MH/25/05/26/6**Date of Verification:** 11-04-2026I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**Gail gas limited**, Address- **Plot No 73, Gail Bhavan, Near Sector 15, Road No 3, CBD Belapur, Navi Mumbai - 400614, Thane, Maharashtra, 400614, Ph:- 7718801142**

| Type of Gas Meter | Manufacturer / Model / Brand / Series | Serial Number | Year of Manufacture | Accuracy Class | Gas Type | Maximum Flow Rate (Qmax) | Minimum Flow Rate (Qmin) | Transitional Flow Rate (Qt) | Size (DN) | G-Value |
|-------------------|---------------------------------------|---------------|---------------------|----------------|-------------|--------------------------|--------------------------|-----------------------------|-----------|---------|
| Diaphragm | GRECO | 20260405360 | 2026 | 1.5 | Natural Gas | 2.5 m3/hr | 0.016 m3/hr | 0.25 m3/hr | 20 | G 1.6 |

| Unit of Measurement (m ³ / kg) | Indicating Device Resolution | Maximum Permissible Error (MPE) | Visual Examination | Accuracy Test at Qmin | Accuracy Test at Qt | Accuracy Test at Qmax | Overall Performance Result |
|---|------------------------------|---------------------------------|--------------------|-----------------------|---------------------|-----------------------|----------------------------|
| m3/hr | 0.01 m3/hr | ±1.5 | Pass | Pass | Pass | Pass | Pass |

| Ambient Temperature | Gas Temperature | Operating Pressure | Relative Humidity | Verification Seal Affixed | Seal Identification No(s) | Software / Firmware Version & Checksum (if applicable) | Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial / billing transactions | Remarks (if any) |
|---------------------|-----------------|--------------------|-------------------|---------------------------|---------------------------|--|--|---|--------------------------|
| 21.4 | 21 | 0.5 bar | 49 | Yes | 3492 | 4.0.1 | Yes | Yes | Meter Found Satisfactory |

Model Approval No(s) - TCM/143/23-5931**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. 590** deposited vide Money Receipt No. **184894646859**, Dated **10-04-2026**Repaired by / Used by: **Green Globe Fuel Solutions Pvt. Ltd.**Next verification falls due on or before: **2033-04-10****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.