

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/05)

Certificate of Verification**Green Globe Fuel Solutions Pvt. Ltd., Green Globe Fuel Solutions Pvt. Ltd. Greenhouse I, II & III, Building No. 5 & 6, Radhe Krishna Industrial Park, Pimplas Village, Mumbai Nashik Road, Bhiwandi, Thane - 421302, Maharashtra- India-421302, Ph: 9820056837****Certificate No:** IND/GATC/MH/25/05/26/7**Date of Verification:** 11-04-2026I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**Gail gas limited**, Address- **Plot No 73, Gail Bhavan, Near Sector 15, Road No 3, CBD Belapur, Navi Mumbai - 400614, Thane, Maharashtra, 400614, Ph:- 7718801142**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	GRECO	20260405361	2026	1.5	Natural Gas	2.5 m3/hr	0.016 m3/hr	0.25 m3/hr	20	G 1.6

Unit of Measurement (m ³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m3/hr	0.01 m3/hr	±1.5	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
21.4	21	0.5 bar	49	Yes	3492	4.0.1	Yes	Yes	Meter Found Satisfactory

Model Approval No(s) - TCM/143/23-5931**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. 590** deposited vide Money Receipt No. **184894646859**, Dated **10-04-2026**Repaired by / Used by: **Green Globe Fuel Solutions Pvt. Ltd.**Next verification falls due on or before: **2033-04-10****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.