

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/RJ/25/03)

Certificate of Verification**Yadav Measurements Private Limited, 373-375, RIICO BHAMA SHAH INDUSTRIAL AREA, KALADWAS, UDAIPUR, Udaipur, Rajasthan, 313001-313003, Ph: 8690222058****Certificate No:** IND/GATC/RJ/25/03/26/25**Date of Verification:** 11-04-2026**I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-HCG DISTRIBUTION PRIVATE LIMITED,Address-HCG DISTRIBUTION PRIVATE LIMITED KHASRA NO.-55/19/1, VILLAGE - DHANKOT, (NEAR SHYAM CHOWK DHANKOT), GURGAON, HARYANA, 122505, IN,Gurugram,Haryana,122505, Ph:- 9990 595 595**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Other	Secure Meters Limited / Liberty Gas 120	Q1964686	2026	1.5	Natural Gas	2.5 m ³ /h	0.016 m ³ /h	0.25 m ³ /h	20	G-1.6

Unit of Measurement (m ³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m ³	0.001	± 2.0 Qmax and Qt±3.5 Qmin	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
-10 to 55	-10 to 55	500 mbar	< 65%	Yes	702490	Metering FW 23420101Metering FW CRC 590C2C4A	Yes	Yes	

Model Approval No(s) - IND/09/24/039**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 590** deposited vide Money Receipt No. **213**, Dated **10-04-2026**Repaired by / Used by: **Yadav Measurements Private Limited**Next verification falls due on or before: **2032-04-10****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.