

Government Approved Test Center Details

**GOVERNMENT APPROVED TEST CENTRE**  
(IND/GATC/TN/25/02)

**Certificate of Verification**  
Government Approved Test Centre  
Ishbaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011332

Certificate No: IND/GATC/TN/25/02/26/01 Date of Verification: 07-02-2026

I hereby certify that I have this day verified and stamped/checked the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg), etc. belonging to M/s: ISHBAA SCALES INC., Add: No. 13- MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE, Coimbatore, Tamil Nadu, 641006, Ph:- 9843014400

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (SI)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (g)	Unit of Measurement	Actual Scale Interval (g)	No. of Verification Intervals (n = Max / g)	Maximum Permissible Error (MPE)
Electronic	ISHBAA	C-1009300	2025	M	50	100	5	kg	5	10000	7.5g

Visual Examination	Zero Setting / Zero Tracking Test	Excentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature 20±5°C	Relative Humidity 40±2%	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	27	44	230v-50Hz	Yes	GATC/TN/25/02/26/01	V.2.0

Instrument conforms to OIML Recommendation L1 (SI) Rules, 2011	Yes	Verified and stamped for use in commercial transactions	Yes	Remarks (if any):
--	-----	---	-----	-------------------

Model Approval No(s) - IND/09/15/609  
 Verification Fee (Rs. + GST) - 2360. Carriage / Conveyance/adjusting Charges, etc. (+ GST) - 0  
 Total Rs. NR deposited vide money Receipt. No.251 Dated 07-02-2026 Repaired/ Used by - Ishbaa Scales Inc.  
 Next verification falls due on or before: 2027-02-06

Note:  
 1. The above-mentioned weights and measures are verified and stamped:  
 1. in the premises of GATC  
 2. in-situ at the place of user  
 2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

Signature of Principal Officer




Department of Consumer Affairs Version 1.0