



# GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/10)

## Certificate of Verification

Government Approved Test Centre

**Pietro Fiorentini DB India Private Limited, Gat No. 43, Navlakh Umbre, Talegaon MIDC,  
Taluka Maval, Pune - 410507 , Maharashtra India.-410507, Ph: 9503304767**

**Certificate No:** IND/GATC/MH/25/10/26/04

**Date of Verification:** 11-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-**Pietro Fiorentini DB India Pvt Ltd.,Address-Gat No. 43, Navlakh Umbre, Behind Talegaon MIDC, Taluka Maval Pune. 410507 Maharashtra India,Pune,Maharashtra,410507, Ph:- 7987546325**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)
Diaphragm	PietroFiorentiniG1.6	260000082	2026	1.5	Natural Gas	2.5 m3/hr	0.016 m3/hr	0.25 m3/hr

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m3	0.0002	1.5 %	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011
20.3	20.3	937.5 mbar (Pabs)	63.2	Yes	XXXXXX	NA	Yes

**Model Approval No(s) - IND/09/24/536**

**Verification Fee (Rs. + GST) - 590. Carriage / Conveyance,adjusting Charges,etc. (+ GST)- 0**

**Total Rs.590 deposited vide money Receipt. No.Pending Dated 11-02-2026 Repaired by/ Used by :  
Pietro Fiorentini DB India Private Limited**

**Next verification falls due on or before: 2033-02-10**

**Note:**

**1.The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**

