

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/RJ/25/06)

**Certificate of Verification****Capital Innotech Pvt., E-100, Mewar Industrial Area, Madri-313003, Ph: 9891079083****Certificate No:** IND/GATC/RJ/25/06/26/14139**Date of Verification:** 18-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**AAVANTIKA GAS LIMITED**, Address-**2nd Floor, 202-B, NRK Business Park, Vijay Nagar Square, A.B Road, Indore, Indore, Madhya Pradesh, 452010, Ph:- 7314222520**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	CAPITAL G1.6MS	20260439983	2026	1.5	Natural Gas	2.5	0.016	0.25	20	1.6

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m <sup>3</sup>	0.0001	Qmax. to Qt ±1.5%, Qt to Qmin. ± 3%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
24±2	24±2	94±3kPa (Pabs)	55±10	Yes	604399836	-	Yes	Yes	-

**Model Approval No(s) - IND/09/22/924****Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **Not Applicable**, Dated **18-04-2026**Repaired by / Used by: **Capital Innotech Pvt.**Next verification falls due on or before: **2033-04-17****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**