

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/HR/26/20)

Certificate of Verification**Precision Electronic Instruments Co., 1256, M.I.E, Part-B, Bahadurgarh, Jhajjar, Haryana - 124507-124507, Ph: 9810004327****Certificate No:** IND/GATC/HR/26/20/26/167**Date of Verification:** 30-04-2026I hereby certify that I have this day verified and stamped/rejected the under mentioned Tape Measures,etc. belonging to M/s-**Precision Electronic Instruments Co.,Address-1256, M.I.E, Part-B, Bahadurgarh, Jhajjar, Haryana,Jhajjar,Haryana,124507, Ph:- 9810004327**

| Instrument Type | Manufacturer & Model / Brand/ Series | Serial Number | Nominal Length | Width | Accuracy Class (I / II / III): | Year of Manufacture |
|-----------------|--------------------------------------|---------------|----------------|-------|--------------------------------|---------------------|
| Steel | Goldtech-GTSM | 26121154 | 200cm | 5.5mm | II | 2026 |

| Unit of Measurement: metre (m) | Scale Interval | Reference Temperature: 20 °C | Maximum Permissible Error (as per OIML / General Rules) |
|--------------------------------|----------------|------------------------------|---|
| 2m | 1mm | No | Rules |

| Visual Examination | Nominal Length Accuracy Test | Scale Interval Accuracy Test | End Hook / Zero Error Check | Overall Accuracy Result | Ambient Temperature: | Relative Humidity | Verification Mark / Stamp Affixed | Seal Identification No(s): |
|--------------------|------------------------------|------------------------------|-----------------------------|-------------------------|----------------------|-------------------|-----------------------------------|----------------------------|
| Pass | Pass | 1mm | Pass | Pass | Yes | Yes | Yes | NA |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped | Remarks (if any) |
|--|----------------------|------------------|
| Yes | Yes | NA |

Model Approval No(s) - IND/09/18/71**Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. N/A** deposited vide Money Receipt No. **NA**, Dated **30-04-2026**Repaired by / Used by: **Precision Electronic Instruments Co.**Next verification falls due on or before: **2028-04-29****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.