

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/25/02)

Certificate of Verification

Government Approved Test Centre

**Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA,
COIMBATORE-641006, Ph: 9843011232**

Certificate No: IND/GATC/TN/25/02/26/26

Date of Verification: 12-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-ISHTAA SCALES INC,Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement
Electronic	ISHTAA-IHB	VGB5023	2026	III	50kg	200g	10g	g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity
Pass	Pass	Pass	Pass	Pass	Pass	Pass	28.9	40

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/09/258

Verification Fee (Rs. + GST) - 3540. Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs.0 deposited vide money Receipt. No.75 Dated 12-02-2026 Repaired by/ Used by : Ishtaa Scales Inc.

Next verification falls due on or before: 2027-02-11

Signature of Principal Officer

Note:

1.The above-mentioned weights and measures are verified and stamped:

i. In the premises of GATC

ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

