

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/10)

Certificate of Verification**Pietro Fiorentini DB India Private Limited, Gat No. 43, Navlakh Umbre, Talegaon MIDC, Taluka Maval, Pune - 410507, Maharashtra India.-410507, Ph: 9503304767****Certificate No:** IND/GATC/MH/25/10/26/1704**Date of Verification:** 05-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-**Maharashtra Natural Gas Limited,Address-M/s. Maharashtra Natural Gas Limited Pot No-27, A-Block, 1st Floor, Narveer Tanajiwadi, PMT Bus Depot Commercial Building First Floor Shivaji nagar Pune-411005,Pune,Maharashtra,411005, Ph:- 9503304767**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	PietrofiorentiniG1.6	260002883	2026	1.5	Natural Gas	2.5 m ³ /hr	0.016 m ³ /hr	0.25 m ³ /hr	20	G1.6

Unit of Measurement (m ³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m ³	0.0002	1.5%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
22.6	22.6	937 mbar (Pabs)	57	Yes	2603P9	NA	Yes	Yes	OK

Model Approval No(s) - IND/09/24/536**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. -, Dated **N/A**Repaired by / Used by: **Pietro Fiorentini DB India Private Limited**Next verification falls due on or before: **2033-04-04****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.