

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/UP/25/11)

Certificate of Verification**CIVA Industrial Solutions, D-177 First Floor, Sector -10, Noida-201301, Ph: 9958050808****Certificate No:** IND/GATC/UP/25/11/26/7**Date of Verification:** 26-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**SOMETHING DIFFERENT RESTAURANTS PVT. LTD.**,Address-**IIIRD FLOOR PLOT NO-003, C-417, DLF MALL SECTOR-18, NOIDA,Gautam Buddha Nagar,Uttar Pradesh,201301, Ph:- 8527126192**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	EAGLE/ECON BS44	DB4047	2023	III	150kg	400g	20g	kg	10g	7500	30g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	26.6	60	220V-50HZ	Yes	IND/GATC//UP/25/11	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	0

Model Approval No(s) - IND/09/16/413**Verification Fee (Rs. + GST) - 3540; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 3540** deposited vide Money Receipt No. **CIS/25-26/00466**, Dated **26-03-2026**Repaired by / Used by: **CIVA Industrial Solutions**Next verification falls due on or before: **2027-03-25****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.