

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/25/02)

Certificate of Verification

Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011232

Certificate No: IND/GATC/TN/25/02/26/394

Date of Verification: 11-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-ISHTAA SCALES INC,Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400

| Type of Instrument | Manufacturer / Model / Brand / Series Designation | Serial Number | Year of Manufacture | Accuracy Class (III) | Maximum Capacity (Max upto 150 kg) | Minimum Capacity (Min) | Verification Scale Interval (e) | Unit of Measurement:kg | Actual Scale Interval (d) | No. of Verification Intervals (n = Max / e) | Maximum Permissible Error (MPE) |
|--------------------|---|---------------|---------------------|----------------------|------------------------------------|------------------------|---------------------------------|------------------------|---------------------------|---|---------------------------------|
| Electronic | ISHTAA-IFB | VGB5390 | 2026 | III | 40/100kg | 100g | 5/10g | kg | 5/10g | 10000 | 15g |

| Visual Examination | Zero Setting / Zero Tracking Test | Eccentricity Test | Repeatability Test | Accuracy / Weighing Performance Test | Tare Device Test | Overall Verification Result | Ambient Temperature | Relative Humidity | Supply Voltage (if electronic) | Verification on Seal Affixed | Seal Identification No(s) | Software Identification / Version /Checksum (if applicable) |
|--------------------|-----------------------------------|-------------------|--------------------|--------------------------------------|------------------|-----------------------------|---------------------|-------------------|--------------------------------|------------------------------|---------------------------|---|
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | 33.1 | 34 | 230V-50HZ | Yes | IND/GATC/TN/25/02 | |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial transactions | Remarks (if any): |
|--|---|-------------------|
| Yes | Yes | |

Model Approval No(s) - IND/09/15/609

Verification Fee (Rs. + GST) - 35400; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. 146, Dated 11-03-2026

Repaired by / Used by: Ishtaa Scales Inc.

Next verification falls due on or before: 2027-03-10

Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

i. In the premises of GATC

ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

