

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/KL/26/04)

Certificate of Verification

Interweighing Pvt. Ltd, 49/470 D1,3rd floor, Asian Tower,Vyttila ,Cochin, Kerala -682019-682019, Ph: 8590601636

Certificate No: IND/GATC/KL/26/04/26/354

Date of Verification: 25-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-MEEZAN ELECTRONIC SCALES PVT LTD,Address-BRIGHT BUILDING, CALICUT ROAD, MANJERI,Malappuram, Kerala,676121, Ph:- 8111854411

Table with 12 columns: Type of Instrument, Manufacturer / Model / Brand / Series Designation, Serial Number, Year of Manufacture, Accuracy Class (III), Maximum Capacity (Max upto 150 kg), Minimum Capacity (Min), Verification Scale Interval (e), Unit of Measurement:kg, Actual Scale Interval (d), No. of Verification Intervals (n = Max / e), Maximum Permissible Error (MPE)

Table with 13 columns: Visual Examination, Zero Setting / Zero Tracking Test, Eccentricity Test, Repeatability Test, Accuracy / Weighing Performance Test, Tare Device Test, Overall Verification Result, Ambient Temperature, Relative Humidity, Supply Voltage (if electronic), Verification on Seal Affixed, Seal Identification No(s), Software Identification / Version /Checksum (if applicable)

Table with 3 columns: Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011, Verified and stamped for use in commercial transactions, Remarks (if any):

Model Approval No(s) - IND/09/21/124

Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A

Total Rs. N/A deposited vide Money Receipt No. N/A, Dated N/A

Repaired by / Used by: Interweighing Pvt. Ltd

Next verification falls due on or before: 2027-04-24



Handwritten signature of the Principal Officer

Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

