

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/PB/25/01)

Certificate of Verification**FMI LIMITED, G.T. ROAD, DORAHA-141421, Ph: 7009989823****Certificate No:** IND/GATC/PB/25/01/26/196**Date of Verification:** 22-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Tape Measures,etc. belonging to M/s-**FMI LIMITED**,Address-**G.T. ROAD, DORAHA,Ludhiana,Punjab,141421, Ph:- 7009989823**


Instrument Type	Manufacturer & Model / Brand/ Series	Serial Number	Nominal Length	Width	Accuracy Class (I / II / III):	Year of Manufacture
Steel	FMI LIMITED	196	3 m	N/A	II	2026

Unit of Measurement: metre (m)	Scale Interval	Reference Temperature: 20 °C	Maximum Permissible Error (as per OIML / General Rules)
metre (m)	1 mm	Yes	Rules

Visual Examination	Nominal Length Accuracy Test	Scale Interval Accuracy Test	End Hook / Zero Error Check	Overall Accuracy Result	Ambient Temperature:	Relative Humidity	Verification Mark / Stamp Affixed	Seal Identification No(s):
Pass	Pass	1 mm	Pass	Pass	Yes	Yes	Yes	N/A

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped	Remarks (if any)
Yes	Yes	

Model Approval No(s) - IND/09/98/219**Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. 0, Dated 22-04-2026Repaired by / Used by: **FMI LIMITED**Next verification falls due on or before: **2028-04-21**


Signature of Principal Officer

Note:**1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.