

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/25/02)

Certificate of Verification

Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011252

Certificate No: IND/GATC/TN/25/02/26/229

Date of Verification: 27-02-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg), etc. belonging to M/s-ISHTAA SCALES INC, Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA, Coimbatore, Tamil Nadu, 641006, Ph:- 9843014400

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement: kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ISHTAA-IPB	VG85226	2026	III	40/100kg	100g	5/10g	kg	5.10g	10000	1.5g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version / Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31.2	32	230V-50HZ	Yes	IND/GATC/TN/25/02	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/09/85

Verification Fee (Rs. + GST) - 7080; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. 125, Dated 27-02-2026

Repaired by / Used by: **Ishtaa Scales Inc.**

Next verification falls due on or before: **2027-02-26**

V. Brindha
Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

