

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/RJ/25/03)

**Certificate of Verification****Yadav Measurements Private Limited, 373-375, RIICO BHAMA SHAH INDUSTRIAL AREA, KALADWAS, UDAIPUR, Udaipur, Rajasthan, 313001-313003, Ph: 8690222058****Certificate No:** IND/GATC/RJ/25/03/26/7**Date of Verification:** 02-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**Yadav Measurements Pvt. Ltd.**, Address-**F-378 to 380, G-421 to 432 G1-390 to 396, H-381 to 389, RIICO - Bhamashah Industrial Area, Kaladwas, Udaipur, Udaipur, Rajasthan, 313003, Ph:- 1234567890**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	YMPL	123456	2026	1.5	Natural Gas	2.5	0.016	0.25	20	G-1.6

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m <sup>3</sup>	0.001	1.5	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
22±2	22±2	20	<60	Yes	123456		Yes	Yes	

**Model Approval No(s) - W&M****Verification Fee (Rs. + GST) - 500; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 90****Total Rs. 590** deposited vide Money Receipt No. **123**, Dated **02-04-2026**Repaired by / Used by: **Yadav Measurements Private Limited**Next verification falls due on or before: **2032-04-01****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**