## GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/25/02)

### Certificate of Verification

**Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011232**

**Certificate No:** IND/GATC/TN/25/02/26/451 **Date of Verification:** 16-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**ISHTAA SCALES INC,**Address-**13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400**

| Type of Instrument | Manufacturer / Model / Brand / Series Designation | Serial Number | Year of Manufacture | Accuracy Class (III) | Maximum Capacity (Max upto 150 kg) | Minimum Capacity (Min) | Verification Scale Interval (e) | Unit of Measurement:kg | Actual Scale Interval (d) | No. of Verification Intervals (n = Max / e) | Maximum Permissible Error (MPE) |
|---|---|---|---|---|---|---|---|---|---|---|---|
| Electronic | ISHTAA-ITP | VGC5447 | 2026 | III | 30kg | 100g | 5g | kg | 5g | 6000 | 7.5g |

| Visual Examination | Zero Setting / Zero Tracking Test | Eccentricity Test | Repeatability Test | Accuracy / Weighing Performance Test | Tare Device Test | Overall Verification Result | Ambient Temperature | Relative Humidity | Supply Voltage (if electronic) | Verification Seal Affixed | Seal Identification No(s) | Software Identification / Version /Checksum (if applicable) |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | 26 | 42 | 230V-50HZ | Yes | IND/GATC/TN/25/02 | |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial transactions | Remarks (if any): |
|---|---|---|
| Yes | Yes | |

**Model Approval No(s) - IND/09/11/163**

**Verification Fee (Rs. + GST) - 17700; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0**

**Total Rs. 0** deposited vide Money Receipt No. **153**, Dated **16-03-2026**
Repaired by / Used by: **Ishtaa Scales Inc.**
Next verification falls due on or before: **2027-03-15**

**Signature of Principal Officer**

**Note:**

**1. The above-mentioned weights and measures are verified and stamped:**

☑ **i. In the premises of GATC**

☐ **ii. In-situ at the place of User**

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**