

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/RJ/25/06)

Certificate of Verification**Capital Innotech Pvt., E-100, Mewar Industrial Area, Madri-313003, Ph: 9891079083****Certificate No:** IND/GATC/RJ/25/06/26/1515**Date of Verification:** 06-04-2026**I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-CUGL,Address-7th Floor, UPSIDC Complex | A-1/4, Lakhapur, Kanpur,Kanpur Nagar,Uttar Pradesh,208024, Ph:- 8308816774**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	CAPITAL G1.6MS	CI-26-02-12496	2026	1.5	Natural Gas	2.5	0.016	2.5	20	G1.6

Unit of Measurement (m ³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
M3	0.01	Qmax to Qt - +/- 1.5%, Qt to Qmin. +/-3%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
-25 to +60	-25 to +60	500mbar	90	Yes	C6300001		Yes	Yes	

Model Approval No(s) - IND/09/22/924**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **Not Applicable**, Dated **06-04-2026**Repaired by / Used by: **Capital Innotech Pvt.**Next verification falls due on or before: **2033-04-06****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.