

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/TN/25/02)

**Certificate of Verification**

**Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011232**

Certificate No: IND/GATC/TN/25/02/26/545

Date of Verification: 23-03-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-ISHTAA SCALES INC,Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ISHTAA-IFB	VGC5541	2026	III	50kg	100g	5g	kg	5g	10000	7.5g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31.1	34	230V-50HZ	Yes	IND/GATC/TN/25/02	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	


Model Approval No(s) - IND/09/15/609

Verification Fee (Rs. + GST) - 35400; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. 172, Dated 23-03-2026

Repaired by / Used by: Ishtaa Scales Inc.

Next verification falls due on or before: 2027-03-22



Signature of Principal Officer

**Note:**

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

