

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/28)

Certificate of Verification

Janhavi Calibration Services LLP, S 87 MIDC-425003, Ph: 9423915430

Certificate No: IND/GATC/MH/26/28/26/23

Date of Verification: 03-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-SSP AGRO,Address-GAT NO.620/2 CHAHARDI,TAL-CHOPDA,Jalgaon,Maharashtra,425434, Ph:- 9764748865

Table with 12 columns: Type of Instrument, Manufacturer / Model / Brand / Series Designation, Serial Number, Year of Manufacture, Accuracy Class (III), Maximum Capacity (Max upto 150 kg), Minimum Capacity (Min), Verification Scale Interval (e), Unit of Measurement:kg, Actual Scale Interval (d), No. of Verification Intervals (n = Max / e), Maximum Permissible Error (MPE). Row 1: Electronic, ATCOM, 200880, III, 30/60 kg, 100 g, 5/10 g, kg, 6000, 15.

Table with 13 columns: Visual Examination, Zero Setting / Zero Tracking Test, Eccentricity Test, Repeatability Test, Accuracy / Weighing Performance Test, Tare Device Test, Overall Verification on Result, Ambient Temperature, Relative Humidity, Supply Voltage (if electronic), Verification on Seal Affixed, Seal Identification No(s), Software Identification / Version /Checksum (if applicable). Row 1: Pass, Pass, Pass, Pass, Pass, Pass, Pass, 32, 46, 230 V, Yes, IND/GATC/MH/26/28.

Table with 3 columns: Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011, Verified and stamped for use in commercial transactions, Remarks (if any). Row 1: Yes, Yes.

Model Approval No(s) - IND/09/07/191

Verification Fee (Rs. + GST) - 708; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 118

Total Rs. 826 deposited vide Money Receipt No. 6, Dated 03-06-2026

Repaired by / Used by: Janhavi Calibration Services LLP

Next verification falls due on or before: 2027-06-02

Sanjeev Shankar Patil, Digitally signed by Sanjeev Shankar Patil, Date: 2026.06.04 07:27:27 +05'30'

Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

