

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/TN/25/02)

**Certificate of Verification**

**Ishtaa Scales Inc., 13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVA INDIA, COIMBATORE-641006, Ph: 9843011232**

**Certificate No: IND/GATC/TN/25/02/26/1237**

**Date of Verification: 29-05-2026**

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to **M/s-ISHTAA SCALES INC,Address-13, MJ SARGUNAM COLONY, NEAR SNR COLLEGE, NAVAINDIA,Coimbatore,Tamil Nadu,641006, Ph:- 9843014400**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ISHTAA-ITB	VGE6232	2026	III	8/20kg	20g	1/2g	kg	1/2g	10000	3g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	28.2	32	230V-50HZ	Yes	IND/GATC/TN/25/02	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

**Model Approval No(s) - IND/09/09/84**

**Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0**

**Total Rs. 0 deposited vide Money Receipt No. N/A, Dated 29-05-2026**

**Repaired by / Used by: Ishtaa Scales Inc.**

**Next verification falls due on or before: 2027-05-28**



**Signature of Principal Officer**

**Note:**

**1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring Instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**

