

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/26/06)

**Certificate of Verification****Sansui Electronics, Survey No 32, Near Pune Weighbridge, Pisoli Road, Pune, Maharashtra-411048, Ph: 9350077677****Certificate No:** IND/GATC/MH/26/06/26/41**Date of Verification:** 03-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**SANSUI ELECTRONICS.**,Address-**PISOLI,NEAR AMBEKAR HOTEL,PUNE-411060,Pune,Maharashtra,411060, Ph:-9823888888**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	SPP	S260680159	2026	III	30/100 kg.	100 g.	5/10 g.	kg.	5/10 g	6000/10000	15 g.

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	35	38	230 V 50 HZ	Yes	IND/GATC/MH/26/06	V 100

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	OK

**Model Approval No(s) - IND/09/94/84****Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. N/A** deposited vide Money Receipt No. **N/A**, Dated **N/A**Repaired by / Used by: **Sansui Electronics**Next verification falls due on or before: **2027-06-02****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**