

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/26/09)

**Certificate of Verification****Weighcal Metrology Services Private Limited, Plot No. BH-1/23, Thakare Nagar, N-2 CIDCO, Aurangabad, Pin code-431001, Ph: 9637700799****Date of Verification: 29-05-2026****Certificate No: IND/GATC/MH/26/09/26/15**

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-Prince Hardware,Address-Nawabpura , Mondha Road,Chhatrapati Sambhajinagar,Maharashtra,431001, Ph:-9860839400

| Type of Instrument | Manufacturer / Model / Brand / Series Designation | Serial Number | Year of Manufacture | Accuracy Class (III) | Maximum Capacity (Max upto 150 kg) | Minimum Capacity (Min) | Verification Scale Interval (e) | Unit of Measurement:kg | Actual Scale Interval (d) | No. of Verification Intervals (n = Max / e) | Maximum Permissible Error (MPE) |
|--------------------|---|---------------|---------------------|----------------------|------------------------------------|------------------------|---------------------------------|------------------------|---------------------------|---|---------------------------------|
| Electronic         | Jarawala Digital Scale                            | 201254        |                     | III                  | 100 Kg                             | 200 g                  | 10 g                            | Kg                     | 10 g                      | 10000                                       | 15 g                            |

| Visual Examination | Zero Setting / Zero Tracking Test | Eccentricity Test | Repeatability Test | Accuracy / Weighing Performance Test | Tare Device Test | Overall Verification Result | Ambient Temperature | Relative Humidity | Supply Voltage (if electronic) | Verification on Seal Affixed | Seal Identification No(s) | Software Identification / Version /Checksum (if applicable) |
|--------------------|-----------------------------------|-------------------|--------------------|--------------------------------------|------------------|-----------------------------|---------------------|-------------------|--------------------------------|------------------------------|---------------------------|---|
| Pass               | Pass                              | Pass              | Pass               | Pass                                 | Pass             | Pass                        | 31                  | 42                | 230v, 50Hz                     | Yes                          | IND/GATC/MH/26/09         |   |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial transactions | Remarks (if any): |
|--|---|-------------------|
| Yes  | Yes   |                   |

**Model Approval No(s) - IND/09/14/408****Verification Fee (Rs. + GST) - 472; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. 472 deposited vide Money Receipt No. 0010, Dated 07-05-2026****Repaired by / Used by: Weighcal Metrology Services Private Limited****Next verification falls due on or before: 2027-05-28**


**Signature of Principal Officer**

**Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**