

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/06)

Certificate of Verification**Sansui Electronics, Survey No 32, Near Pune Weighbridge, Pisoli Road, Pune, Maharashtra-411048, Ph: 9350077677****Certificate No:** IND/GATC/MH/26/06/26/40**Date of Verification:** 03-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**SANSUI ELECTRONICS.**,Address-**PISOLI,NEAR AMBEKAR HOTEL,PUNE-411060,Pune,Maharashtra,411060, Ph:-9823888888**

| Type of Instrument | Manufacturer / Model / Brand / Series Designation | Serial Number | Year of Manufacture | Accuracy Class (III) | Maximum Capacity (Max upto 150 kg) | Minimum Capacity (Min) | Verification Scale Interval (e) | Unit of Measurement:kg | Actual Scale Interval (d) | No. of Verification Intervals (n = Max / e) | Maximum Permissible Error (MPE) |
|--------------------|---|---------------|---------------------|----------------------|------------------------------------|------------------------|---------------------------------|------------------------|---------------------------|---|---------------------------------|
| Electronic | SPP | S260680158 | 2026 | III | 30/100 kg. | 100 g. | 5/10 g. | kg. | 5/10 g | 6000/10000 | 15 g. |

| Visual Examination | Zero Setting / Zero Tracking Test | Eccentricity Test | Repeatability Test | Accuracy / Weighing Performance Test | Tare Device Test | Overall Verification Result | Ambient Temperature | Relative Humidity | Supply Voltage (if electronic) | Verification on Seal Affixed | Seal Identification No(s) | Software Identification / Version /Checksum (if applicable) |
|--------------------|-----------------------------------|-------------------|--------------------|--------------------------------------|------------------|-----------------------------|---------------------|-------------------|--------------------------------|------------------------------|---------------------------|---|
| Pass | Pass | Pass | Pass | Pass | Pass | Pass | 35 | 38 | 230 V 50 HZ | Yes | IND/GATC/MH/26/06 | V 100 |

| Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial transactions | Remarks (if any): |
|--|---|-------------------|
| Yes | Yes | OK |

Model Approval No(s) - IND/09/94/84**Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. N/A** deposited vide Money Receipt No. **N/A**, Dated **N/A**Repaired by / Used by: **Sansui Electronics**Next verification falls due on or before: **2027-06-02****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.