GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/26/16)

Certificate of Verification

ASIAN SCALE INDUSTRIES / PROPRIETOR GOBIVASAN A, 174-B1, LAKSHMIAPURAM STREET, (BALAJI PRESS OPP) PACR SALAI, RAJAPALAYAM-626117, Ph: 8675718383

Certificate No: IND/GATC/TN/26/16/26/12 **Date of Verification:** 01-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**ASIAN TECH,**Address-**174 - B1,LAKSHMIAPURAM STREET, P.A.C.R SALAI, RAJAPALAYAM,Virudhunagar,Tamil Nadu,626117, Ph:- 8675718383**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ASIAN TECH	CS-200005	2026	III	40/80Kg	100g	5/10g	Kg	5/10g	8000	15g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	34.5	42	230AC±10 %50HZ	Yes	IND/GATC/ TN/16/26	NIL

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	ORIGINAL VERIFICATION

Model Approval No(s) - IND/09/19/325

Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. **0**, Dated **01-06-2026**
Repaired by / Used by: **ASIAN SCALE INDUSTRIES / PROPRIETOR GOBIVASAN A**
Next verification falls due on or before: **2027-05-31**

Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

☑ **i. In the premises of GATC**

☐ **ii. In-situ at the place of User**

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.