

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/HR/26/07)

Certificate of Verification

Saini Scales, Plot No.2 Saini Colony, Jhansa Road, Opp. Jindal Park, Kurukshetra, Haryana-136118, Ph: 9017541000

Certificate No: IND/GATC/HR/26/07/26/1

Date of Verification: 31-05-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg).etc. belonging to M/s-**JETHU MAL SURAJ BHAN**,Address-NGM, **SHOP NO. 115, KAITHAL,Kaithal,Haryana,136027, Ph:- 9416383117**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	CASTEC	26BP7337	2026	III	150KG	2KG	100GM	KG	100GM	1500	150GM

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification on Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	30.4	45	220	Yes	IND/GATC/26/07/HAR/26 B	SAINI SCALES V-6

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	YES	

Model Approval No(s) - IND/09/16/05

Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A

Total Rs. N/A deposited vide Money Receipt No. **N/A**, Dated **N/A**

Repaired by / Used by: **Saini Scales**

Next verification falls due on or before: **2027-05-30**

Ahanna

Signature of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

