

# GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/26/26)

## Certificate of Verification

Government Approved Test Centre

**SALEM LEGAL METROLOGY LABORATORIES LLP, 9/172-E, Thamarai Garden, Near Thamarai Nagar, Kondappanaickenpatty, Kannankurichy Post, Salem - 636 008.-636008, Ph: 9790034962**

**Certificate No:** IND/GATC/TN/26/26/2

**Date of Verification:** 02-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**M.THILAGARAJ,Address-82, MURUGAN MALAIGAI, ANNA NAGAR Ist CROSS, PONNAMMAPET,Salem,Tamil Nadu,636001, Ph:- 9600580486**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ESSAE / DS-252 / G 25207148767	G 25207148767	2004	III	10/20 kg	20 g	1 g / 2 g	kg	1 g / 2 g	10000	6 g


Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	34	54	190 to 270 V	Yes	IND/GATC/TN/26/26	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	NA

**Model Approval No(s) - IND/09/2004/245**

**Verification Fee (Rs. + GST) - 590. Carriage / Conveyance,adjusting Charges,etc. (+ GST)- 0**

**Total Rs.590 deposited vide money Receipt. No.SLML/GATC/26/06/B002 Dated 02-06-2026 Repaired by/ Used by : SALEM LEGAL METROLOGY LABORATORIES LLP**  
**Next verification falls due on or before: 2027-06-01**

  
Signature of Principal Officer

**Note:**

**1.The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**

