# GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/10)

## Certificate of Verification

**Pietro Fiorentini DB India Private Limited, Gat No. 43, Navlakh Umbre, Talegaon MIDC, Taluka Maval, Pune - 410507 , Maharashtra India.-410507, Ph: 9503304767**

**Certificate No:** IND/GATC/MH/25/10/26/4550 **Date of Verification:** 14-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-**Aavantika Gas Limited,** Address-**202-B,NRK Business Park,Vijay Nagar Square,Indore-Indore 452010, Madhya Pradesh India,Indore,Madhya Pradesh,452010, Ph:- 9503304767**

| Type of Gas Meter | Manufacturer / Model / Brand/ Series | Serial Number | Year of Manufacture | Accuracy Class | Gas Type | Maximum Flow Rate (Qmax) | Minimum Flow Rate (Qmin) | Transitional Flow Rate (Qt) | Size (DN) | G-Value |
|---|---|---|---|---|---|---|---|---|---|---|
| Diaphragm | PietrofiorentiniG1.6 | 260005752 | 2026 | 1.5 | Natural Gas | 2.5 m³/hr | 0.016 m³/hr | 0.25 m³/hr | 20 | G1.6 |

| Unit of Measurement (m³ / kg) | Indicating Device Resolution | Maximum Permissible Error (MPE) | Visual Examination | Accuracy Test at Qmin | Accuracy Test at Qt | Accuracy Test at Qmax | Overall Performance Result |
|---|---|---|---|---|---|---|---|
| m³ | 0.0002 | 1.5% | Pass | Pass | Pass | Pass | Pass |

| Ambient Temperature | Gas Temperature | Operating Pressure | Relative Humidity | Verification Seal Affixed | Seal Identification No(s) | Software / Firmware Version & Checksum (if applicable) | Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011 | Verified and stamped for use in commercial / billing transactions | Remarks (if any) |
|---|---|---|---|---|---|---|---|---|---|
| 22.6 | 22.6 | 936 mbar (Pabs) | 55 | Yes | AGL2604P5 | NA | Yes | Yes | OK |

**Model Approval No(s) - IND/09/24/536**

**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0**

**Total Rs. 0** deposited vide Money Receipt No. **-**, Dated **N/A**
Repaired by / Used by: **Pietro Fiorentini DB India Private Limited**
Next verification falls due on or before: **2033-04-13**

Digitally signed by SATHISH KADIRVELU
Date: 2026.05.11 10:51:31 +05:30
Location: Talegoan

**Signature of Principal Officer**

**Note:**

**1. The above-mentioned weights and measures are verified and stamped:**

☑ **i. In the premises of GATC**

☐ **ii. In-situ at the place of User**

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**