

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/26/28)

**Certificate of Verification****Janhavi Calibration Services LLP, S 87 MIDC-425003, Ph: 9423915430****Certificate No: IND/GATC/MH/26/28/26/19****Date of Verification: 03-06-2026**

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**SSP AGRO,Address-GAT NO.620/2 CHAHARDI,TAL-CHOPDA,Jalgaon,Maharashtra,425434, Ph:- 9764748865**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ATCOM	200608		III	30/60 kg	100 g	5/10 g	kg		6000	15

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	32	46	230 V	Yes	IND/GATC/MH/26/28	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

**Model Approval No(s) - IND/09/07/191****Verification Fee (Rs. + GST) - 708; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 118****Total Rs. 826** deposited vide Money Receipt No. 6, Dated **03-06-2026**Repaired by / Used by: **Janhavi Calibration Services LLP**Next verification falls due on or before: **2027-06-02**

Sanjeev  
Shankar Patil

Digitally signed by  
Sanjeev Shankar Patil  
Date: 2026.06.04  
07:22:32 +05'30'

**Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**