

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/25/10)

**Certificate of Verification****Pietro Fiorentini DB India Private Limited, Gat No. 43, Navlakh Umbre, Talegaon MIDC, Taluka Maval, Pune - 410507, Maharashtra India.-410507, Ph: 9503304767****Certificate No:** IND/GATC/MH/25/10/26/6269**Date of Verification:** 18-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters, etc. belonging to M/s-**Aavantika Gas Limited**, Address-**202-B, NRK Business Park, Vijay Nagar Square, Indore-Indore 452010, Madhya Pradesh India, Indore, Madhya Pradesh, 452010, Ph:- 9503304767**

Type of Gas Meter	Manufacturer / Model / Brand / Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	Pietrofiorentini G1.6	260005368	2026	1.5	Natural Gas	2.5 m <sup>3</sup> /hr	0.016 m <sup>3</sup> /hr	0.25 m <sup>3</sup> /hr	20	G1.6

Unit of Measurement (m <sup>3</sup> / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m <sup>3</sup>	0.0002	1.5%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
22.6	22.6	936 mbar (Pabs)	58	Yes	AGL2604P3	NA	Yes	Yes	OK

**Model Approval No(s) - IND/09/24/536****Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. -, Dated **N/A**Repaired by / Used by: **Pietro Fiorentini DB India Private Limited**Next verification falls due on or before: **2033-04-17****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**