

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/UP/26/01)

**Certificate of Verification****BHASKAR SCALE ENGINEERING, B398 PATEL NAGAR 2ND GHAZIABAD-201001, Ph: 9871254243****Certificate No:** IND/GATC/UP/26/01/26/73**Date of Verification:** 03-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-M/s **Bhaskar Scale Engineering**,Address-B **398, PATEL NAGAR 2ND, GHAZIABAD, UTTAR PRADESH,Ghaziabad, Uttar Pradesh,201001, Ph:- 9871254243**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	BHASKAR	BTT2026273	2026	Class III	20kg	40g	2g	kg	2g	10000	3g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	25	50%	240V 50Hz	Yes	INDGATCU P2601B	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	101

**Model Approval No(s) - IND/09/19/428****Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **0**, Dated **03-06-2026**Repaired by / Used by: **BHASKAR SCALE ENGINEERING**Next verification falls due on or before: **2027-06-02****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**