

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/26/26)

Certificate of Verification

Government Approved Test Centre

SALEM LEGAL METROLOGY LABORATORIES LLP, 9/172-E, Thamarai Garden, Near Thamarai Nagar, Kondappanaickenpatty, Kannankurichy Post,
Salem - 636 008.-636008, Ph: 9790034962

Certificate No: IND/GATC/TN/26/26/1

Date of Verification: 02-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-M/S MUTHUGANESH BHARATGAS AGENCY,Address-No.70 Salai Road, Ponnammampet, Salem - 636001.,Salem,Tamil Nadu,636001, Ph:- 9442506276

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ISHTAA SCALES INC / IHB 50	C-1016713	2024	III	50 kg	200 g	10 g	kg	10 g	5000	10 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	25.5	44	230 V - 50 Hz	Yes	IND/GATC/TN/26/26	1.0

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	NA

Model Approval No(s) - IND/GATC/TN/26/26

Verification Fee (Rs. + GST) - 590. Carriage / Conveyance,adjusting Charges,etc. (+ GST)- 0

Total Rs.590 deposited vide money Receipt. No.SLML/GATC/26/06/B001 Dated 02-06-2026 Repaired by/ Used by : SALEM LEGAL METROLOGY LABORATORIES LLP
Next verification falls due on or before: 2027-06-01


Signature of Principal Officer

Note:

1.The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

