

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/KL/26/04)

**Certificate of Verification****Interweighing Pvt. Ltd, 49/470 D1,3rd floor, Asian Tower,Vyttila ,Cochin, Kerala -682019-682019, Ph: 8590601636****Certificate No:** IND/GATC/KL/26/04/26/1106**Date of Verification:** 02-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**INTERWEIGHING PVT LTD,Address-49/470 D1 3RD FLOOR,ASIAN TOWER,VYTTILA,COCHIN,KERALA-682019, Ernakulam,Kerala,682019, Ph:- 8590601636**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	YESWEIGH	Y09708	2026	III	30kg	100g	5g	kg	5g	6000	7.5g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification on Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	29.3	73	230 V AC	Yes	IND/KL/26/04/B26	Nil

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	tested ok

**Model Approval No(s) - IND/09/20/23****Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. N/A** deposited vide Money Receipt No. **N/A**, Dated **N/A**Repaired by / Used by: **Interweighing Pvt. Ltd**Next verification falls due on or before: **2027-06-01****HARISH  
RAMANKUTTY**Digitally signed by  
**HARISH RAMANKUTTY**  
Date: 2026.06.02  
10:44:54 +05'30'**Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC  
 ii. In-situ at the place of User

**2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.**