

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/28)

Certificate of Verification

Janhavi Calibration Services LLP, S 87 MIDC-425003, Ph: 9423915430

Certificate No: IND/GATC/MH/26/28/26/30

Date of Verification: 04-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**ATCOM ENTERPRISES,Address-S-87 MIDC JALGAON,Jalgaon,Maharashtra,425003, Ph:- 9420607083**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ATCOM	200883	2026	III	100 kg	200 g	10 g	kg		10000	15 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification on Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31	45	230 V	Yes	IND/GATC/MH/26/28	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/07/191

Verification Fee (Rs. + GST) - 472; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A

Total Rs. 472 deposited vide Money Receipt No. **8**, Dated **04-06-2026**

Repaired by / Used by: **Janhavi Calibration Services LLP**

Next verification falls due on or before: **2027-06-03**

Sanjeev
Shankar Patil

Digitally signed by
Sanjeev Shankar Patil
Date: 2026.06.04
11:46:34 +05'30'

Signature of Principal Officer



Note:

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

