

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/28)

Certificate of Verification**Janhavi Calibration Services LLP, S 87 MIDC-425003, Ph: 9423915430****Certificate No:** IND/GATC/MH/26/28/26/38**Date of Verification:** 05-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**BALAJI PLASTOTECH**,Address-**V-110/A, V-110/B MIDC, JALGAON,Jalgaon,Maharashtra,425003, Ph:- 9225800312**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	ATCOM	200886	2026	III	50/150 kg	200 g	10/20 g	kg		7500	30 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31	46	230 V	Yes	IND/GATC/MH/26/28	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/07/191**Verification Fee (Rs. + GST) - 472; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. 472** deposited vide Money Receipt No. **10**, Dated **05-06-2026**Repaired by / Used by: **Janhavi Calibration Services LLP**Next verification falls due on or before: **2027-06-04****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.