

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/TN/26/15)

Certificate of Verification**ALEX ZACHARIAH JOHN, 2/1132, THANTHAI PERIYAR NAGAR 7TH STREET ,BURMA COLONY KARAUKUDI-630002, Ph: 9443111148****Certificate No:** IND/GATC/TN/26/15/26/51**Date of Verification:** 15-07-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**A. KADHAR ALI ,ONION SALES ,VEGETABLE MARKET,Address-DEVAKOTTAI MARKET,Sivaganga,Tamil Nadu, 630302, Ph:- 9360264794**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	LAKSHMI	2507218	2025	III	20kg	40g	2g	kilogram	2g	10000	3g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	36	85	220	Yes	IND/GATC/TN/2615	NIL

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	NIL

Model Approval No(s) - IND/09/22/982LSS-TT**Verification Fee (Rs. + GST) - 1186; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 59****Total Rs. 1245** deposited vide Money Receipt No. **0034**, Dated **12-07-2026**Repaired by / Used by: **ALEX ZACHARIAH JOHN**Next verification falls due on or before: **2027-07-14****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- ☐ i. In the premises of GATC
- ☒ ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.