

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/09)

Certificate of Verification**Weighcal Metrology Services Private Limited, Plot No. BH-1/23, Thakare Nagar, N-2 CIDCO, Aurangabad, Pin code-431001, Ph: 9637700799****Certificate No: IND/GATC/MH/26/09/26/21****Date of Verification: 29-05-2026**I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to **M/s-A R HARDWARE,Address-Opp. VIP Function Hall , Central Naka,Chhatrapati Sambhajinagar,Maharashtra, 434100, Ph:- N/A**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	LU. Raja Gordhan & CO.	JE 4614		III	20 Kg	40 g	2 g	Kg	2 g	10000	3 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31	41	230 v , 50 Hz	Yes	IND/GATC/MH/26/09	

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	

Model Approval No(s) - IND/09/11/260**Verification Fee (Rs. + GST) - 354; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A****Total Rs. 354 deposited vide Money Receipt No. 0016, Dated 21-05-2026****Repaired by / Used by: ~~Weighcal Metrology Services Private Limited~~****Next verification falls due on or before: 2027-05-28**


Signature of Principal Officer

Note:**1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.