

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/26/27)

Certificate of Verification**Nitiraj Engineers Limited (Calibration Division), J25 & J26, M.I.D.C.,Awdhan , Dhule-424006, Ph: 9158243400****Certificate No:** IND/GATC/MH/26/27/26/587**Date of Verification:** 15-07-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**Nitiraj Engineers Limited (Calibration Division)**,Address-**J25 and J26, M.I.D.C. Awdhan,Dhule,Maharashtra, 424006, Ph:- 9158243400**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	NPW-8/20/30	M600776	2026	III	8/20/30 kg	20 g	1/2/5 g	kg	1/2/5 g	8000/10000/6000	7.5 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	29	42	230V-50Hz	Yes	IND/GATC/MH/26/27	NA

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	NA

Model Approval No(s) - IND/09/15/505**Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **NA**, Dated **15-07-2026**Repaired by / Used by: **Nitiraj Engineers Limited (Calibration Division)**Next verification falls due on or before: **2027-07-14****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- ☒ i. In the premises of GATC
- ☐ ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

