

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/HR/26/21)

Certificate of Verification**Precision Electronic Instruments Co., 1256, M.I.E, Part-B, Bahadurgarh, Jhajjar, Haryana - 124507-124507, Ph: 9810004327****Certificate No:** IND/GATC/HR/26/21/26/3744**Date of Verification:** 15-06-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-**Precision Electronic Instruments Co.,Address-1256, M.I.E, Part-B, Bahadurgarh, Jhajjar, Haryana,Jhajjar, Haryana,124507, Ph:- 9810004327**

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	Goldtech-GTEP	26226189	2026	III	120 kg	2 kg	100 gm	kg	100 gm	1200	100 gm

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification on Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	31.2	34	4v	Yes	HR 26 21	Oxca

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	No	New goods checked by Lokesh Yadav

Model Approval No(s) - IND/09/18/623**Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **0**, Dated **N/A**Repaired by / Used by: **Precision Electronic Instruments Co.**Next verification falls due on or before: **2027-06-14****Signature valid**Digitally signed by **RAJNISH KUMAR AGGARWAL**
Date: 2026.07.06 16:11:42 +05'30'**Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- ☒ i. In the premises of GATC
- ☐ ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.