

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/MH/25/10)

Certificate of Verification

Pietro Fiorentini DB India Private Limited, Gat No. 43, Navlakh Umbre, Talegaon MIDC, Taluka Maval, Pune - 410507 , Maharashtra India.-410507, Ph: 9503304767

Certificate No: IND/GATC/MH/25/10/26/16448**Date of Verification:** 25-04-2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned Gas meters,etc. belonging to M/s-**Aavantika Gas Limited, Address-202-B,NRK Business Park,Vijay Nagar Square,Indore-Indore 452010, Madhya Pradesh India,Indore,Madhya Pradesh,452010, Ph:- 9503304767**

Type of Gas Meter	Manufacturer / Model / Brand/ Series	Serial Number	Year of Manufacture	Accuracy Class	Gas Type	Maximum Flow Rate (Qmax)	Minimum Flow Rate (Qmin)	Transitional Flow Rate (Qt)	Size (DN)	G-Value
Diaphragm	Pietrofiorentini G1.6	260012802	2026	1.5	Natural Gas	2.5 m³/hr	0.016 m³/hr	0.25 m³/hr	20	G1.6

Unit of Measurement (m³ / kg)	Indicating Device Resolution	Maximum Permissible Error (MPE)	Visual Examination	Accuracy Test at Qmin	Accuracy Test at Qt	Accuracy Test at Qmax	Overall Performance Result
m³	0.0002	1.5%	Pass	Pass	Pass	Pass	Pass

Ambient Temperature	Gas Temperature	Operating Pressure	Relative Humidity	Verification Seal Affixed	Seal Identification No(s)	Software / Firmware Version & Checksum (if applicable)	Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial / billing transactions	Remarks (if any)
22.6	22.6	936 mbar(Pabs)	58	Yes	AGL2604P41	NA	Yes	Yes	OK

Model Approval No(s) - IND/09/24/536**Verification Fee (Rs. + GST) - 590; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. -, Dated **N/A**Repaired by / Used by: **Pietro Fiorentini DB India Private Limited**Next verification falls due on or before: **2033-04-24****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- ☒ i. In the premises of GATC
- ☐ ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

