

**GOVERNMENT APPROVED TEST CENTRE**

(IND/GATC/MH/26/06)

**Certificate of Verification**

**Sansui Electronics, Survey No 32, Near Pune Weighbridge, Pisoli Road, Pune, Maharashtra-411048, Ph: 9350077677**

**Certificate No: IND/GATC/MH/26/06/26/9**

**Date of Verification: 01-06-2026**

I hereby certify that I have this day verified and stamped/rejected the under mentioned Non-automatic weighing instruments of Accuracy Class III (upto 150kg),etc. belonging to M/s-SANSUI ELECTRONICS,Address-NEAR AMBEKAR HOTEL , PISOLI , PUNE 411060,Pune,Maharashtra,411060, Ph:-9823888888

Type of Instrument	Manufacturer / Model / Brand / Series Designation	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement:kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Electronic	SSP	S260560359	2026	III	20 kg	40 g	2 g	kg	2 g	10000	3 g

Visual Examination	Zero Setting / Zero Tracking Test	Eccentricity Test	Repeatability Test	Accuracy / Weighing Performance Test	Tare Device Test	Overall Verification Result	Ambient Temperature	Relative Humidity	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version /Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	30	35	230 V 50 Hz	Yes	IND/GATC/MH/26/06	v 100

Instrument conforms to OIML Recommendation/ LM (Gen) Rules, 2011	Verified and stamped for use in commercial transactions	Remarks (if any):
Yes	Yes	OK

**Model Approval No(s) - IND/09/2002/06**

**Verification Fee (Rs. + GST) - N/A; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - N/A**

**Total Rs. N/A deposited vide Money Receipt No. N/A, Dated N/A**

**Repaired by / Used by: Sansui Electronics**

**Next verification falls due on or before: 2027-05-31**



*[Handwritten Signature]*

Signature of Principal Officer

**Note:**

1. The above-mentioned weights and measures are verified and stamped:

- i. In the premises of GATC
- ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.

