

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/PB/25/01)

Certificate of Verification**FMI LIMITED, G.T. ROAD, DORAHA-141421, Ph: 7009989823****Certificate No:** IND/GATC/PB/25/01/26/792**Date of Verification:** 12-05-2026


I hereby certify that I have this day verified and stamped/rejected the under mentioned Tape Measures,etc. belonging to M/s-**FMI LIMITED**,Address-**G.T. ROAD, DORAHA,Ludhiana,Punjab,141421, Ph:- 7009989823**

Instrument Type	Manufacturer & Model / Brand / Series	Serial Number	Nominal Length	Width	Accuracy Class (I / II / III):	Year of Manufacture
Steel	FMI LIMITED	792	3m	N/A	II	2026

Unit of Measurement: metre (m)	Scale Interval	Reference Temperature: 20 °C	Maximum Permissible Error (as per OIML / General Rules)
metre (m)	1mm	Yes	Rules

Visual Examination	Nominal Length Accuracy Test	Scale Interval Accuracy Test	End Hook / Zero Error Check	Overall Accuracy Result	Ambient Temperature:	Relative Humidity	Verification Mark / Stamp Affixed	Seal Identification No(s):
Pass	Pass	1mm	Pass	Pass	Yes	Yes	Yes	N/A

Instrument conforms to OIML Recommendation / LM (Gen) Rules, 2011	Verified and stamped	Remarks (if any)
Yes	Yes	.

Model Approval No(s) - IND/09/98/219**Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0****Total Rs. 0** deposited vide Money Receipt No. **0**, Dated **12-05-2026**Repaired by / Used by: **FMI LIMITED**Next verification falls due on or before: **2028-05-11****Signature of Principal Officer****Note:****1. The above-mentioned weights and measures are verified and stamped:**

- i. In the premises of GATC
 ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.